

VMI - system





Fig 2A

BEST AVAILABLE COPY



The beam after the two Acusto Optics Modulators




Fig 2C




Fig 3





Fig 6

BEST AVAILABLE COPY



— Experiment Theory

Fig 5

FO2001" E5102650




Fig 4

BEST AVAILABLE COPY

ChDiv

Input - two vectors: $Y(i)$ - channel 1 - square wave - chopping signal, $0 \leq Y_i \leq 1$

$X(i)$ - channel 2 - fluorescence raw data - from the detecting region (both line scan)


Usually Sampled at 40KHz

Output - two vectors: $OutRight(j)$ - fluorescence from Right Line scan

$OutLeft(j)$ - fluorescence from Left Line scan

Usually Sampled at 5KHz

The sampling rate of the output channels always equals the frequency of the chopping signal



* Program ends when input file ends

Fig 7

BEST AVAILABLE COPY


ArVlAnalyzer

Input: two files (one for each line scan).

Each file contain 2 vectors one of Positions ($P(i)$) and the other has the corresponding Area ($A(i)$)

Output: three vectors - Area, TimeDiff (inversely proportional to velocity), Position

Parameters that can be determined - MinTimeDiff, MaxTimeDiff



Position is presented in point number and not time

TimeDiff is in Seconds and is inversely proportional to the velocity

* Program ends when one of the input files ends

Fig 8

BEST AVAILABLE COPY